

<b>Notice of Allowability</b>	<b>Application No.</b>	<b>Applicant(s)</b>	
	10/719,028	NAKAGAWA, MASAO	
	Examiner Tilahun B. Gesesse	Art Unit 2618	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1.  This communication is responsive to 11/05/07.
  2.  The allowed claim(s) is/are 1-4.
  3.  Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
    - a)  All    b)  Some\*    c)  None    of the:
      1.  Certified copies of the priority documents have been received.
      2.  Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
      3.  Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).
- \* Certified copies not received: \_\_\_\_\_.
- Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.  
**THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.**
4.  A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
  5.  CORRECTED DRAWINGS ( as "replacement sheets") must be submitted.
    - (a)  including changes required by the Notice of Draftsperson's Patent Drawing Review ( PTO-948) attached
      - 1)  hereto or 2)  to Paper No./Mail Date \_\_\_\_\_.
    - (b)  including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date \_\_\_\_\_.

Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
  6.  DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

#### Attachment(s)

1.  Notice of References Cited (PTO-892)
2.  Notice of Draftsperson's Patent Drawing Review (PTO-948)
3.  Information Disclosure Statements (PTO/SB/08),  
Paper No./Mail Date \_\_\_\_\_
4.  Examiner's Comment Regarding Requirement for Deposit  
of Biological Material
5.  Notice of Informal Patent Application
6.  Interview Summary (PTO-413),  
Paper No./Mail Date 1/15/08
7.  Examiner's Amendment/Comment
8.  Examiner's Statement of Reasons for Allowance
9.  Other \_\_\_\_\_.

1. An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with Mr. William Brooks on January 15, 2008.

The application has been amended as follows:

In claim 1, line 1, after the term UWB --- repeater--- has been inserted.

In claim 1, line 4, after the word a relay controller---arranged between said receiver and transmitter---has been inserted.

In claim 1, line 5, after the word receiver ---said relay controller detecting receive timing of the UWB pulse series signals and establishing transmit timing so that the transmit timing does not overlap the receiving timing signals, said relay controller transmitting said pulse series signals---has been inserted.

In claim 4, line 1, after the word UWB ---repeater---has been inserted.

In claim 4, line 4, after the word controller --- arranged between said receiver and transmitter for detecting receive timing of the UWB pulse series and setting ---has been inserted.

In claim 4, line 4 after the word controller "sets" has been deleted.

In claim 4, line 7, after transmitter--- and such that the transit timing does not overlap the receive timing signals--- has been added.

***Allowable Subject Matter***

2. The following is an examiner's statement of reasons for allowance:

The prior art of record does not teach a UWB repeater transmitter and receiver, transmits at a predetermined transmitting timing which is different from the receiving timing by a pulse unit in order to prevent a sneak path wave between the receiver and the transmitter. These limitation, in conjunction with all other limitations, has not been disclosed or suggested or made obvious over the prior art of record.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

***Conclusion***

3. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Tilahun B. Gesesse whose telephone number is 571-272-7879. The examiner can normally be reached on flex.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Matthew Anderson can be reached on 571-272-4177. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

January 14, 2008

  
TILAHUN GESESSE  
PRIMARY EXAMINER

Tilahun B. Gesesse  
Primary Examiner  
Art Unit 2618